
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000242
10:33 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:12
STERILIZING 4:49
136.0 C 223kPa 4:49
136.4 C 233kPa 5:19
136.4 C 233kPa 5:49
136.4 C 233kPa 6:19
136.4 C 233kPa 6:49
136.4 C 233kPa 7:19
136.3 C 232kPa 7:49
136.1 C 231kPa 8:19
136.1 C 231kPa 8:19
VENTING 8:19
AIR DRYING 9:23
CYCLE COMPLETE 23:24

Digital Signature #
B242FEA78C7D8B05
