
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001815
10:48 08/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:44
STERILIZING 3:44
136.0 C 220kPa 3:44
136.2 C 223kPa 4:14
136.3 C 224kPa 4:44
136.4 C 225kPa 5:14
136.4 C 225kPa 5:44
136.4 C 226kPa 6:14
136.1 C 222kPa 6:44
136.4 C 225kPa 7:14
136.4 C 225kPa 7:14
VENTING 7:14
AIR DRYING 8:06
CYCLE COMPLETE 38:06

Digital Signature #
9ABA8BCD571D52AD
