
Midmark M11 - v1.0.0
Total Cycles: 5094

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2024 05:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.8 F 0.0
2:00 135.4 F 0.1
4:00 161.3 F 0.6
6:00 182.5 F 1.2
8:00 199.8 F 4.9
10:00 216.0 F 8.7
12:00 227.7 F 13.0
14:00 232.3 F 8.7
16:00 241.1 F 11.7
18:00 243.4 F 12.0
20:00 241.2 F 10.9
22:00 254.5 F 17.4
24:00 265.5 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.3 F 28.1
3:00 272.4 F 28.7
4:00 271.4 F 28.0

Min 271.3 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 209.4 F 0.0
10:00 197.0 F 0.0
15:00 195.7 F 0.0
20:00 210.8 F 0.0
25:00 228.6 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 25:48
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:30
CYCLE COMPLETE

11/14/2024 06:42 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____