

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2107

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2022 02:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
2:00 183.5 F 4.0  
4:00 206.2 F 9.0  
6:00 221.0 F 8.7  
8:00 231.0 F 9.1  
10:00 237.0 F 9.7  
12:00 243.1 F 12.1  
14:00 240.2 F 10.4  
16:00 253.9 F 17.2  
18:00 265.5 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 272.7 F 28.9  
2:00 271.7 F 28.1  
3:00 272.7 F 29.0  
4:00 272.3 F 28.4  
  
Min 271.6 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.7 F 0.0  
5:00 202.4 F 0.0  
10:00 200.8 F 0.0  
15:00 209.8 F 0.0  
20:00 178.0 F 0.0  
25:00 174.2 F 0.0  
30:00 175.4 F 0.0  
Drying Complete

FILLING: 2:22  
HEATING: 19:47  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

07/08/2022 03:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_