
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000472
8:47 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:03
STERILIZING 4:05
136.5 C 220kPa 4:05
136.3 C 223kPa 4:35
136.4 C 224kPa 5:05
136.4 C 224kPa 5:35
136.4 C 224kPa 6:05
136.4 C 221kPa 6:35
136.4 C 219kPa 7:05
136.3 C 223kPa 7:35
136.3 C 223kPa 7:35
VENTING 7:35
AIR DRYING 8:28
CYCLE COMPLETE 8:33

Digital Signature #
608DCE803C142855
