
Midmark M11 - v1.0.0
Total Cycles: 4091

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 174.8 F 2.9
4:00 195.2 F 7.4
6:00 212.8 F 12.3
8:00 224.0 F 6.4
10:00 237.5 F 11.5
12:00 234.0 F 8.5
14:00 241.9 F 10.8
16:00 246.8 F 13.4
18:00 259.0 F 19.9
20:00 269.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.8 F 28.4
2:00 272.4 F 28.5
3:00 272.4 F 28.8
4:00 272.4 F 29.0

Min 271.5 F 28.3
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 193.0 F 0.0
10:00 212.6 F 0.0
15:00 204.2 F 0.0
20:00 194.2 F 0.0
25:00 192.2 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

12/06/2023 02:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____