
Midmark M11 - v1.0.0
Total Cycles: 3943

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2023 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.9 F 0.0
2:00 123.6 F 0.2
4:00 140.2 F 2.9
6:00 174.1 F 6.6
8:00 197.5 F 11.2
10:00 218.0 F 4.7
12:00 227.8 F 9.2
14:00 236.7 F 10.9
16:00 237.6 F 10.3
18:00 238.1 F 8.2
20:00 248.4 F 14.5
22:00 260.5 F 21.2
24:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 29.0
2:00 271.3 F 28.3
3:00 272.6 F 29.1
4:00 271.8 F 28.5

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 213.1 F 0.0
10:00 217.1 F 0.0
15:00 218.0 F 0.0
20:00 217.8 F 0.0
25:00 215.6 F 0.0
30:00 213.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:56
CYCLE COMPLETE

10/24/2023 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____