

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2530

Name: RITTER2  
Sterilizer ID: V2313228  
06/19/2023 11:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.3 F 0.0  
2:00 186.8 F 3.7  
4:00 205.4 F 9.3  
6:00 224.1 F 7.9  
8:00 233.4 F 11.8  
10:00 236.7 F 9.8  
12:00 236.6 F 8.8  
14:00 249.7 F 15.4  
16:00 261.7 F 22.4  
18:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 272.7 F 29.8  
2:00 271.9 F 29.4  
3:00 272.7 F 30.0  
4:00 272.1 F 29.3  
  
Min 271.8 F 29.2  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.4 F 0.0  
5:00 230.0 F 0.0  
10:00 221.6 F 0.0  
15:00 227.9 F 0.0  
20:00 226.7 F 0.0  
25:00 225.9 F 0.0  
30:00 224.0 F 0.0  
Drying Complete

FILLING: 2:19  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

06/19/2023 12:50 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_