
Midmark M11 - v1.0.0
Total Cycles: 4322

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2024 12:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.2 F 0.0
2:00 116.1 F 1.1
4:00 129.7 F 3.3
6:00 156.3 F 6.1
8:00 181.6 F 9.4
10:00 196.5 F 13.5
12:00 215.2 F 5.5
14:00 228.2 F 9.3
16:00 238.0 F 12.4
18:00 235.7 F 9.9
20:00 240.5 F 9.4
22:00 245.0 F 13.2
24:00 255.9 F 18.9
26:00 265.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.5
2:00 272.2 F 29.0
3:00 271.1 F 28.6
4:00 272.0 F 29.4

Min 271.1 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 221.2 F 0.0
10:00 229.3 F 0.0
15:00 228.3 F 0.0
20:00 223.7 F 0.0
25:00 224.9 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:52
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:39
CYCLE COMPLETE

02/22/2024 02:03 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____