

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002367  
12:42 06/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:54  
PRESSURIZING 2:03  
STERILIZING 3:09  
136.4 C 226kPa 3:09  
136.3 C 225kPa 3:39  
136.3 C 225kPa 4:09  
136.1 C 223kPa 4:39  
136.4 C 226kPa 5:09  
136.4 C 224kPa 5:39  
136.4 C 226kPa 6:09  
136.2 C 226kPa 6:39  
136.2 C 226kPa 6:39  
VENTING 6:39  
AIR DRYING 7:30  
CYCLE COMPLETE 37:30  
-----

Digital Signature #  
03CD0EAD8FE88B84  
-----