
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006921
9:20 09/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:52
STERILIZING 3:55
136.0 C 220kPa 3:55
136.2 C 222kPa 4:25
136.3 C 222kPa 4:55
136.4 C 224kPa 5:25
136.3 C 222kPa 5:55
136.2 C 221kPa 6:25
136.2 C 224kPa 6:55
136.4 C 222kPa 7:25
136.4 C 222kPa 7:25
VENTING 7:25
AIR DRYING 8:14
CYCLE COMPLETE 8:28

Digital Signature #
6627640CEC7F7AC4
