
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001338
10:09 10/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:12
STERILIZING 5:44
136.5 C 224kPa 5:44
136.3 C 222kPa 6:14
136.2 C 222kPa 6:44
136.4 C 224kPa 7:14
136.4 C 223kPa 7:44
136.3 C 223kPa 8:14
136.3 C 224kPa 8:44
136.4 C 225kPa 9:14
136.4 C 225kPa 9:14
VENTING 9:14
AIR DRYING 10:12
CYCLE COMPLETE 15:20

Digital Signature #
FC4BF067B8E0AE09
