
Midmark M11 - v1.0.5
Total Cycles: 2381

Name: RITTER2
Sterilizer ID: V2313228
05/17/2023 07:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.2 F 0.0
2:00 79.4 F 0.0
4:00 107.1 F 0.0
6:00 116.2 F 0.3
8:00 151.2 F 0.4
10:00 177.7 F 2.0
12:00 202.1 F 5.6
14:00 212.3 F 9.5
16:00 224.8 F 13.5
18:00 229.6 F 8.5
20:00 240.9 F 13.3
22:00 239.9 F 11.1
24:00 237.5 F 9.5
26:00 250.9 F 16.1
28:00 262.2 F 22.7
30:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.6 F 29.5
2:00 271.6 F 28.7
3:00 272.7 F 29.7
4:00 272.0 F 29.0

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 192.2 F 0.0
10:00 206.2 F 0.0
15:00 215.7 F 0.0
20:00 213.5 F 0.0
25:00 214.7 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:23
CYCLE COMPLETE

05/17/2023 08:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____