
Midmark M11 - v1.0.0
Total Cycles: 5152

Name: RITTER 1
Sterilizer ID: V2178198
12/03/2024 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.5 F 0.0
2:00 102.1 F 0.0
4:00 134.1 F 0.1
6:00 160.2 F 0.4
8:00 179.7 F 1.1
10:00 198.5 F 4.2
12:00 213.7 F 7.8
14:00 224.0 F 11.9
16:00 228.7 F 7.1
18:00 239.1 F 12.3
20:00 239.0 F 10.1
22:00 237.4 F 8.6
24:00 250.6 F 15.3
26:00 261.8 F 21.7
28:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.6 F 29.0
2:00 272.5 F 28.7
3:00 271.9 F 28.2
4:00 271.9 F 28.6

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 196.7 F 0.0
10:00 189.3 F 0.0
15:00 189.0 F 0.0
20:00 200.0 F 0.0
25:00 207.9 F 0.0
30:00 211.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:42
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:07:16
CYCLE COMPLETE

12/03/2024 04:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____