
Midmark M11 - v1.0.5
Total Cycles: 1712

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 05:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.2 F 0.0
2:00 193.7 F 3.2
4:00 203.1 F 8.9
6:00 221.5 F 8.8
8:00 232.2 F 11.2
10:00 235.9 F 9.5
12:00 237.0 F 9.1
14:00 250.3 F 15.9
16:00 262.6 F 23.1
18:00 272.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.6
1:00 272.5 F 29.7
2:00 272.8 F 30.2
3:00 272.0 F 29.7
4:00 272.6 F 29.9

Min 271.9 F 29.5
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 226.5 F 0.0
10:00 238.9 F 0.0
15:00 234.2 F 0.0
20:00 228.2 F 0.0
25:00 226.7 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

11/10/2022 06:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____