

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2704

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2022 05:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.1 F 0.0  
2:00 192.9 F 4.4  
4:00 201.1 F 9.2  
6:00 221.8 F 6.7  
8:00 233.9 F 10.1  
10:00 236.4 F 7.8  
12:00 244.8 F 12.6  
14:00 245.2 F 12.7  
16:00 258.3 F 19.7  
18:00 270.0 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.3 F 28.7  
2:00 271.9 F 28.7  
3:00 272.2 F 28.6  
4:00 271.7 F 28.6  
  
Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 232.0 F 0.0  
10:00 228.1 F 0.0  
15:00 228.7 F 0.0  
20:00 228.0 F 0.0  
25:00 230.1 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 18:53  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

12/20/2022 06:43 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_