
Midmark M11 - v1.0.5
Total Cycles: 4466

Name: RITTER2
Sterilizer ID: V2313228
03/27/2025 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.7 F 0.0
2:00 131.6 F 1.6
4:00 151.2 F 4.7
6:00 177.5 F 8.5
8:00 195.9 F 12.7
10:00 216.0 F 5.3
12:00 225.2 F 9.4
14:00 232.5 F 12.0
16:00 237.4 F 10.5
18:00 236.6 F 8.7
20:00 247.8 F 14.7
22:00 258.7 F 20.8
24:00 268.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.2 F 30.2
2:00 272.3 F 29.9
3:00 271.5 F 29.5
4:00 272.0 F 29.9

Min 270.9 F 29.2
Max 273.6 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 184.2 F 0.0
10:00 188.9 F 0.0
15:00 195.4 F 0.0
20:00 194.6 F 0.0
25:00 185.6 F 0.0
30:00 179.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:20
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:04:20
CYCLE COMPLETE

03/27/2025 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____