
Midmark M11 - v1.0.0
Total Cycles: 4172

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2024 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.9 F 0.0
2:00 89.2 F 0.3
4:00 79.8 F 1.2
6:00 107.3 F 3.3
8:00 137.4 F 6.0
10:00 167.9 F 9.2
12:00 196.2 F 12.9
14:00 214.0 F 3.8
16:00 224.5 F 7.0
18:00 232.3 F 10.6
20:00 231.6 F 7.8
22:00 238.4 F 10.2
24:00 236.5 F 8.8
26:00 245.6 F 12.9
28:00 256.4 F 18.8
30:00 266.6 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.8 F 29.1
2:00 272.3 F 29.0
3:00 271.6 F 28.4
4:00 272.7 F 28.8

Min 271.0 F 28.0
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.5 F 0.0
5:00 229.2 F 0.0
10:00 215.7 F 0.0
15:00 219.7 F 0.0
20:00 220.6 F 0.0
25:00 230.9 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

01/03/2024 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____