
Midmark M11 - v1.0.0
Total Cycles: 4182

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2024 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.4 F 0.0
2:00 161.2 F 2.6
4:00 179.7 F 6.2
6:00 198.9 F 10.7
8:00 215.1 F 4.0
10:00 226.6 F 7.7
12:00 235.3 F 12.2
14:00 234.9 F 8.7
16:00 243.1 F 11.8
18:00 244.7 F 12.5
20:00 256.0 F 18.4
22:00 266.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.4 F 28.1
3:00 272.4 F 29.0
4:00 272.1 F 28.6

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 213.5 F 0.0
10:00 236.3 F 0.0
15:00 219.1 F 0.0
20:00 214.2 F 0.0
25:00 210.0 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 23:45
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

01/04/2024 03:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____