
Midmark M11 - v1.0.5
Total Cycles: 4199

Name: RITTER2
Sterilizer ID: V2313228
12/10/2024 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.1 F 0.0
2:00 101.1 F 0.1
4:00 112.8 F 1.3
6:00 146.1 F 4.3
8:00 174.1 F 7.6
10:00 193.7 F 11.1
12:00 206.6 F 4.9
14:00 218.6 F 6.8
16:00 229.0 F 10.8
18:00 230.4 F 7.7
20:00 242.4 F 12.6
22:00 240.8 F 11.2
24:00 252.8 F 17.2
26:00 263.5 F 23.5
28:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.2
2:00 272.0 F 29.5
3:00 271.8 F 28.9
4:00 272.3 F 29.7

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 193.3 F 0.0
10:00 187.3 F 0.0
15:00 186.1 F 0.0
20:00 182.7 F 0.0
25:00 183.5 F 0.0
30:00 184.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:11
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:48
CYCLE COMPLETE

12/10/2024 01:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____