
Midmark M11 - v1.0.0
Total Cycles: 2601

Name: RITTER 1
Sterilizer ID: V2178198
11/23/2022 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 173.7 F 2.9
4:00 184.0 F 7.0
6:00 201.7 F 11.9
8:00 219.8 F 5.5
10:00 234.5 F 10.8
12:00 233.9 F 7.2
14:00 244.5 F 13.4
16:00 243.5 F 11.8
18:00 256.3 F 18.7
20:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.2 F 28.9
2:00 271.4 F 28.5
3:00 272.5 F 28.9
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 220.3 F 0.0
10:00 232.4 F 0.0
15:00 229.3 F 0.0
20:00 229.7 F 0.0
25:00 229.9 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 21:17
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

11/23/2022 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____