
Midmark M11 - v1.0.0
Total Cycles: 4966

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2024 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.3 F 0.0
2:00 122.5 F 0.2
4:00 146.0 F 0.9
6:00 176.2 F 2.8
8:00 189.2 F 6.4
10:00 205.2 F 10.6
12:00 223.3 F 5.9
14:00 227.2 F 9.0
16:00 239.7 F 11.9
18:00 239.2 F 10.6
20:00 235.3 F 8.2
22:00 249.3 F 14.7
24:00 260.4 F 20.9
26:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.9
2:00 271.8 F 28.3
3:00 271.5 F 28.4
4:00 272.4 F 28.9

Min 271.3 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 194.9 F 0.0
10:00 194.1 F 0.0
15:00 194.0 F 0.0
20:00 212.1 F 0.0
25:00 217.8 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:05:16
CYCLE COMPLETE

10/04/2024 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____