
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002393
11:21 17/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 4:02
136.0 C 220kPa 4:02
136.0 C 222kPa 4:32
136.4 C 222kPa 5:02
136.4 C 223kPa 5:32
136.4 C 224kPa 6:02
136.2 C 222kPa 6:32
136.4 C 226kPa 7:02
136.1 C 221kPa 7:32
136.1 C 221kPa 7:32
VENTING 7:32
AIR DRYING 8:29
CYCLE COMPLETE 9:03

Digital Signature #
20C5D1B25F28E55B
