
Midmark M11 - v1.0.0
Total Cycles: 749

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2021 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.8 F 0.0
2:00 201.9 F 4.1
4:00 213.7 F 10.4
6:00 223.3 F 6.2
8:00 237.1 F 12.5
10:00 238.6 F 10.4
12:00 236.7 F 8.9
14:00 252.3 F 16.3
16:00 264.6 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.7 F 28.2
2:00 272.2 F 28.6
3:00 272.5 F 28.9
4:00 271.7 F 28.4

Min 271.5 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.9 F 0.0
5:00 230.7 F 0.0
10:00 229.4 F 0.0
15:00 231.2 F 0.0
20:00 227.4 F 0.0
25:00 227.7 F 0.0
Dry Cycle Aborted

FILLING: 2:17
HEATING: 17:46
STERILIZING: 4:00
VENTING: 2:55
DRYING: 28:60
TOTAL CYCLE: 00:55:58

03/25/2021 02:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____