

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3962

Name: RITTER 1  
Sterilizer ID: V2178198  
10/31/2023 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.7 F 0.0  
2:00 117.3 F 1.0  
4:00 137.5 F 3.7  
6:00 167.2 F 7.8  
8:00 190.3 F 13.0  
10:00 212.5 F 5.0  
12:00 226.0 F 9.4  
14:00 233.2 F 9.5  
16:00 239.9 F 11.1  
18:00 237.0 F 8.9  
20:00 251.1 F 15.6  
22:00 263.0 F 22.4  
24:00 271.5 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.0 F 28.2  
2:00 272.7 F 28.7  
3:00 271.7 F 28.2  
4:00 272.1 F 28.2

Min 271.5 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 206.2 F 0.0  
10:00 226.3 F 0.0  
15:00 226.3 F 0.0  
20:00 224.0 F 0.0  
25:00 226.1 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 24:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:02:37  
CYCLE COMPLETE

10/31/2023 12:21 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_