

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001215  
13:44 17/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:51  
PRESSURIZING 3:53  
STERILIZING 5:15  
136.0 C 220kPa 5:15  
136.4 C 225kPa 5:45  
136.2 C 223kPa 6:15  
136.3 C 224kPa 6:45  
136.4 C 225kPa 7:15  
136.3 C 225kPa 7:45  
136.4 C 224kPa 8:15  
136.4 C 224kPa 8:45  
136.4 C 224kPa 8:45  
VENTING 8:45  
AIR DRYING 9:43  
CYCLE COMPLETE 12:23

-----  
Digital Signature #  
9A141AF208955B1A  
-----