

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3800

Name: RITTER 1  
Sterilizer ID: V2178198  
09/20/2023 11:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.1 F 0.0  
2:00 170.1 F 2.6  
4:00 192.2 F 6.7  
6:00 210.9 F 11.4  
8:00 220.7 F 5.8  
10:00 233.3 F 11.0  
12:00 232.5 F 8.0  
14:00 244.2 F 12.2  
16:00 244.5 F 12.5  
18:00 256.2 F 18.7  
20:00 266.8 F 25.1  
22:00 272.5 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.1 F 28.7  
2:00 272.4 F 29.0  
3:00 271.9 F 29.0  
4:00 272.0 F 28.7

Min 271.6 F 28.5  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 257.3 F 0.0  
10:00 231.9 F 0.0  
15:00 228.0 F 0.0  
20:00 225.8 F 0.0  
25:00 232.7 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 22:14  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

09/20/2023 12:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_