
Midmark M11 - v1.0.0
Total Cycles: 3163

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2023 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 170.9 F 0.0
4:00 199.0 F 0.0
6:00 217.4 F 4.8
8:00 230.4 F 9.6
10:00 236.1 F 9.4
12:00 242.1 F 11.9
14:00 240.1 F 10.3
16:00 237.9 F 9.2
18:00 253.6 F 16.8
20:00 266.1 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.2 F 28.3
2:00 272.1 F 28.5
3:00 272.0 F 28.1
4:00 272.1 F 28.5

Min 271.5 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 221.5 F 0.0
10:00 236.9 F 0.0
15:00 231.6 F 0.0
20:00 227.7 F 0.0
25:00 224.7 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

04/25/2023 03:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____