

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4357

Name: RITTER2  
Sterilizer ID: V2313228  
02/19/2025 12:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.1 F 0.0  
2:00 94.4 F 0.4  
4:00 98.0 F 1.3  
6:00 138.8 F 3.6  
8:00 169.2 F 6.5  
10:00 209.2 F 9.9  
12:00 216.9 F 9.8  
14:00 210.7 F 5.3  
16:00 222.3 F 8.8  
18:00 232.0 F 12.7  
20:00 233.6 F 8.8  
22:00 244.2 F 12.9  
24:00 242.5 F 11.9  
26:00 253.0 F 17.3  
28:00 262.3 F 23.1  
30:00 271.5 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 271.9 F 29.9  
2:00 272.4 F 29.6  
3:00 271.0 F 29.4  
4:00 272.6 F 30.4

Min 270.9 F 29.1  
Max 273.1 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 198.2 F 0.0  
10:00 201.7 F 0.0  
15:00 202.7 F 0.0  
20:00 199.0 F 0.0  
25:00 198.9 F 0.0  
30:00 200.3 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 30:33  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:09:23  
CYCLE COMPLETE

02/19/2025 01:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_