

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001598  
9:59 16/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:47  
STERILIZING 3:54  
136.0 C 220kPa 3:54  
136.4 C 225kPa 4:24  
136.4 C 225kPa 4:54  
136.2 C 221kPa 5:24  
136.2 C 223kPa 5:54  
136.4 C 225kPa 6:24  
136.3 C 223kPa 6:54  
136.3 C 224kPa 7:24  
136.3 C 224kPa 7:24  
VENTING 7:24  
AIR DRYING 8:18  
CYCLE COMPLETE 8:26  
-----

Digital Signature #  
628071624992C232  
-----