

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001529  
11:14 10/25/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:37  
PRESSURIZING 6:20  
STERILIZING 7:48  
136.0 C 220kPa 7:48  
136.4 C 217kPa 8:18  
136.2 C 220kPa 8:49  
136.4 C 221kPa 9:19  
136.4 C 218kPa 9:49  
136.2 C 221kPa 10:19  
136.4 C 222kPa 10:49  
136.5 C 222kPa 11:18  
VENTING 11:18  
AIR DRYING 12:23  
CYCLE COMPLETE 42:23  
-----

Digital Signature #  
DBBA558793D6F411  
-----