
Midmark M11 - v1.0.5
Total Cycles: 1794

Name: RITTER2
Sterilizer ID: V2313228
11/30/2022 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 191.3 F 3.6
4:00 206.0 F 9.5
6:00 223.1 F 7.4
8:00 236.0 F 12.1
10:00 238.4 F 10.9
12:00 239.0 F 10.6
14:00 253.8 F 18.1
16:00 266.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.2
1:00 272.7 F 30.0
2:00 271.5 F 29.2
3:00 272.8 F 29.8
4:00 271.6 F 29.1

Min 271.5 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 224.5 F 0.0
10:00 230.4 F 0.0
15:00 229.3 F 0.0
20:00 225.0 F 0.0
25:00 222.3 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 17:21
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:56:12
CYCLE COMPLETE

11/30/2022 03:06 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____