

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1823

Name: RITTER2  
Sterilizer ID: V2313228  
12/08/2022 09:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.5 F 0.0  
2:00 177.7 F 0.0  
4:00 176.1 F 3.1  
6:00 198.4 F 8.2  
8:00 220.1 F 11.6  
10:00 231.2 F 9.4  
12:00 234.3 F 9.3  
14:00 242.9 F 13.1  
16:00 242.8 F 12.2  
18:00 256.1 F 19.1  
20:00 267.6 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.9  
1:00 271.7 F 29.3  
2:00 272.5 F 29.5  
3:00 271.7 F 28.9  
4:00 271.4 F 29.1

Min 271.2 F 28.7  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
5:00 216.1 F 0.0  
10:00 226.6 F 0.0  
15:00 228.6 F 0.0  
20:00 229.2 F 0.0  
25:00 229.1 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:59:10  
CYCLE COMPLETE

12/08/2022 10:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_