

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2631

Name: RITTER2  
Sterilizer ID: V2313228  
07/11/2023 10:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.9 F 0.0  
2:00 142.4 F 0.5  
4:00 152.8 F 3.5  
6:00 183.1 F 7.8  
8:00 204.3 F 12.3  
10:00 217.9 F 6.4  
12:00 230.1 F 11.1  
14:00 230.1 F 7.7  
16:00 242.6 F 13.4  
18:00 242.0 F 12.0  
20:00 254.0 F 18.1  
22:00 264.6 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.5 F 29.9  
2:00 271.2 F 29.2  
3:00 272.5 F 30.1  
4:00 271.7 F 29.4

Min 271.2 F 29.1  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 210.6 F 0.0  
10:00 213.9 F 0.0  
15:00 212.2 F 0.0  
20:00 210.0 F 0.0  
25:00 207.9 F 0.0  
30:00 207.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 23:57  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:02:08  
CYCLE COMPLETE

07/11/2023 11:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_