

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003854  
8:21 22/06/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:49  
PRESSURIZING 2:53  
STERILIZING 3:57  
136.5 C 224kPa 3:57  
136.2 C 222kPa 4:27  
136.4 C 222kPa 4:57  
136.3 C 219kPa 5:27  
136.1 C 222kPa 5:57  
136.2 C 222kPa 6:27  
136.3 C 222kPa 6:57  
136.0 C 220kPa 7:27  
136.0 C 220kPa 7:27  
VENTING 7:27  
AIR DRYING 8:16  
CYCLE COMPLETE 8:40

-----  
Digital Signature #  
5B1AD9DA2E1EE7E9  
-----