
Midmark M11 - v1.0.0
Total Cycles: 4525

Name: RITTER 1
Sterilizer ID: V2178198
05/06/2024 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 68.8 F 0.0
4:00 79.0 F 0.5
6:00 113.4 F 2.1
8:00 141.4 F 4.3
10:00 166.1 F 6.8
12:00 188.7 F 10.1
14:00 206.0 F 10.6
16:00 211.0 F 4.7
18:00 219.9 F 8.2
20:00 230.4 F 11.9
22:00 230.8 F 7.5
24:00 241.9 F 12.6
26:00 239.8 F 10.4
28:00 251.9 F 16.3
30:00 262.3 F 22.3
32:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.1
2:00 271.7 F 28.4
3:00 271.7 F 28.8
4:00 272.4 F 28.9

Min 271.1 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 193.4 F 0.0
10:00 186.9 F 0.0
15:00 186.9 F 0.0
20:00 194.9 F 0.0
25:00 231.0 F 0.0
30:00 228.8 F 0.0

Drying Complete

FILLING: 1:10
HEATING: 32:34
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:11:19
CYCLE COMPLETE

05/06/2024 09:36 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____