
Midmark M11 - v1.0.5
Total Cycles: 1709

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 192.2 F 2.7
4:00 191.6 F 7.6
6:00 211.0 F 13.1
8:00 227.2 F 8.9
10:00 235.9 F 10.0
12:00 242.2 F 13.0
14:00 243.0 F 12.3
16:00 256.6 F 19.4
18:00 268.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 29.9
2:00 271.5 F 29.2
3:00 271.9 F 29.9
4:00 272.3 F 29.8

Min 271.2 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 223.9 F 0.0
10:00 236.7 F 0.0
15:00 228.6 F 0.0
20:00 225.1 F 0.0
25:00 223.8 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

11/10/2022 03:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____