

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000011  
9:31 19/12/2019  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CYCLE ABORTED 1:52  
CYCLE FAULT 15  
NOT STERILE  
END TIME 1:52  
-----

Digital Signature #  
E9F3CF40D012F600  
-----