
Midmark M11 - v1.0.0
Total Cycles: 3828

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2023 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 165.8 F 0.0
4:00 160.2 F 1.2
6:00 192.1 F 4.3
8:00 201.4 F 8.3
10:00 216.9 F 12.6
12:00 224.1 F 7.5
14:00 236.7 F 12.3
16:00 237.6 F 9.6
18:00 239.2 F 9.5
20:00 248.5 F 14.2
22:00 259.6 F 20.3
24:00 269.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.5 F 28.6
2:00 271.5 F 28.3
3:00 272.3 F 28.5
4:00 271.3 F 28.1

Min 271.3 F 28.0
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 222.3 F 0.0
10:00 230.2 F 0.0
15:00 227.9 F 0.0
20:00 228.3 F 0.0
25:00 231.4 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:58
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:26
CYCLE COMPLETE

09/27/2023 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____