
Midmark M11 - v1.0.0
Total Cycles: 5414

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2025 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.4 F 0.0
2:00 158.2 F 3.3
4:00 168.1 F 6.9
6:00 191.4 F 10.8
8:00 207.4 F 5.4
10:00 221.4 F 8.8
12:00 233.9 F 13.4
14:00 234.5 F 9.7
16:00 237.7 F 9.3
18:00 244.6 F 13.2
20:00 256.2 F 19.2
22:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.6
2:00 272.4 F 29.7
3:00 271.9 F 29.7
4:00 272.8 F 29.8

Min 271.0 F 28.8
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 192.9 F 0.0
10:00 214.0 F 0.0
15:00 212.9 F 0.0
20:00 208.7 F 0.0
25:00 212.8 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 23:34
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:24
CYCLE COMPLETE

03/13/2025 05:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____