
Midmark M11 - v1.0.5
Total Cycles: 3939

Name: RITTER2
Sterilizer ID: V2313228
09/12/2024 04:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.4 F 0.0
2:00 165.7 F 2.7
4:00 176.2 F 6.2
6:00 200.5 F 10.1
8:00 216.9 F 7.4
10:00 224.9 F 8.1
12:00 235.6 F 12.8
14:00 236.6 F 10.1
16:00 237.2 F 9.2
18:00 248.2 F 14.7
20:00 259.6 F 20.9
22:00 269.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.6
2:00 271.5 F 28.8
3:00 272.3 F 29.7
4:00 272.0 F 29.3

Min 271.2 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 224.0 F 0.0
10:00 228.7 F 0.0
15:00 202.2 F 0.0
20:00 192.6 F 0.0
25:00 187.3 F 0.0
30:00 184.1 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:02:19
CYCLE COMPLETE

09/12/2024 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____