
Midmark M11 - v1.0.0
Total Cycles: 2403

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2022 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 126.1 F 0.4
4:00 146.5 F 2.1
6:00 183.8 F 5.3
8:00 202.7 F 10.0
10:00 217.6 F 5.9
12:00 225.9 F 8.2
14:00 237.1 F 13.1
16:00 237.6 F 10.0
18:00 236.3 F 8.5
20:00 248.3 F 14.5
22:00 260.5 F 21.2
24:00 271.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.9 F 29.3
2:00 271.9 F 28.5
3:00 272.5 F 29.2
4:00 272.5 F 29.0

Min 271.6 F 28.3
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 221.1 F 0.0
10:00 225.6 F 0.0
15:00 230.1 F 0.0
20:00 231.4 F 0.0
25:00 227.3 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

10/05/2022 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____