
Midmark M11 - v1.0.5
Total Cycles: 1832

Name: RITTER2
Sterilizer ID: V2313228
12/13/2022 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.1 F 0.0
2:00 175.8 F 1.3
4:00 178.4 F 5.6
6:00 204.9 F 10.5
8:00 223.0 F 5.4
10:00 234.0 F 11.4
12:00 234.1 F 8.6
14:00 240.6 F 10.3
16:00 246.7 F 14.0
18:00 259.3 F 20.9
20:00 270.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.6
2:00 271.6 F 28.9
3:00 271.8 F 29.5
4:00 272.5 F 29.5

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 206.3 F 0.0
10:00 220.1 F 0.0
15:00 229.9 F 0.0
20:00 230.8 F 0.0
25:00 227.6 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

12/13/2022 01:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____