

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3090

Name: RITTER 1  
Sterilizer ID: V2178198  
04/11/2023 11:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.6 F 0.0  
2:00 188.3 F 3.9  
4:00 202.3 F 8.6  
6:00 217.9 F 10.5  
8:00 228.8 F 9.2  
10:00 235.8 F 9.4  
12:00 242.8 F 12.6  
14:00 241.8 F 11.2  
16:00 255.1 F 18.1  
18:00 266.9 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.0  
2:00 271.5 F 28.4  
3:00 272.4 F 29.1  
4:00 271.5 F 28.4  
  
Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 224.6 F 0.0  
10:00 238.0 F 0.0  
15:00 227.2 F 0.0  
20:00 225.8 F 0.0  
25:00 224.2 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 19:30  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

04/11/2023 12:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_