
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000833
12:12 31/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 5:28
STERILIZING 6:50
136.0 C 220kPa 6:50
136.4 C 223kPa 7:20
136.3 C 220kPa 7:50
136.3 C 220kPa 8:20
136.4 C 223kPa 8:50
135.8 C 218kPa 9:20
136.2 C 222kPa 9:50
136.4 C 221kPa 10:20
136.4 C 221kPa 10:20
VENTING 10:20
AIR DRYING 11:19
CYCLE COMPLETE 41:19

Digital Signature #
0D495114B1676DD8
