
Midmark M11 - v1.0.5
Total Cycles: 3257

Name: RITTER2
Sterilizer ID: V2313228
12/27/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 157.1 F 2.7
4:00 177.7 F 6.6
6:00 199.8 F 11.9
8:00 222.8 F 6.5
10:00 237.4 F 12.0
12:00 235.8 F 9.4
14:00 237.7 F 9.2
16:00 249.0 F 15.1
18:00 261.8 F 22.3
20:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.3 F 29.2
2:00 272.9 F 29.5
3:00 272.9 F 29.8
4:00 271.8 F 29.5

Min 271.7 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 254.3 F 0.0
10:00 205.3 F 0.0
15:00 195.7 F 0.0
20:00 188.8 F 0.0
25:00 186.1 F 0.0
30:00 184.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

12/27/2023 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____