
Midmark M11 - v1.0.5
Total Cycles: 3286

Name: RITTER2
Sterilizer ID: V2313228
01/10/2024 05:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.2 F 0.0
2:00 128.5 F 0.9
4:00 149.1 F 3.5
6:00 175.1 F 7.2
8:00 194.9 F 11.3
10:00 212.9 F 4.9
12:00 223.8 F 9.3
14:00 232.5 F 11.2
16:00 237.2 F 11.1
18:00 235.1 F 9.3
20:00 248.4 F 15.4
22:00 259.4 F 21.6
24:00 269.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.4
1:00 272.5 F 30.1
2:00 271.7 F 29.5
3:00 271.2 F 29.5
4:00 272.4 F 30.2

Min 271.0 F 29.2
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 233.7 F 0.0
10:00 223.5 F 0.0
15:00 213.4 F 0.0
20:00 226.1 F 0.0
25:00 230.3 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 25:03
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:53
CYCLE COMPLETE

01/10/2024 06:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____