
Midmark M11 - v1.0.0
Total Cycles: 3670

Name: RITTER 1
Sterilizer ID: V2178198
08/17/2023 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.1 F 0.0
2:00 201.3 F 3.5
4:00 210.7 F 9.5
6:00 223.1 F 7.1
8:00 234.2 F 10.2
10:00 234.1 F 7.9
12:00 244.1 F 12.0
14:00 245.7 F 12.9
16:00 258.5 F 19.8
18:00 269.7 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.8 F 28.2
2:00 272.3 F 28.5
3:00 272.7 F 28.9
4:00 271.6 F 28.3

Min 271.6 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 209.6 F 0.0
10:00 217.6 F 0.0
15:00 237.7 F 0.0
20:00 228.8 F 0.0
25:00 226.7 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

08/17/2023 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____