
Midmark M11 - v1.0.0
Total Cycles: 3733

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2023 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.8 F 0.0
2:00 183.8 F 2.1
4:00 200.0 F 6.2
6:00 207.8 F 11.0
8:00 222.2 F 5.7
10:00 234.0 F 10.9
12:00 234.1 F 7.9
14:00 243.0 F 12.2
16:00 244.3 F 12.6
18:00 256.1 F 18.8
20:00 266.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.2
2:00 271.4 F 28.5
3:00 272.1 F 29.3
4:00 272.0 F 29.0

Min 271.0 F 28.4
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 203.4 F 0.0
10:00 215.1 F 0.0
15:00 219.4 F 0.0
20:00 219.6 F 0.0
25:00 223.7 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

09/07/2023 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____