

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000841  
16:10 05/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:17  
STERILIZING 4:27  
136.0 C 228kPa 4:27  
136.1 C 231kPa 4:57  
136.4 C 233kPa 5:27  
136.2 C 232kPa 5:57  
136.3 C 231kPa 6:27  
136.4 C 232kPa 6:57  
136.4 C 233kPa 7:27  
136.4 C 233kPa 7:57  
136.4 C 233kPa 7:57  
VENTING 7:57  
AIR DRYING 8:52  
CYCLE COMPLETE 15:29  
-----

Digital Signature #  
C0455B19054CD6D0  
-----